

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**AND**

**ANSI/NCSL Z540-1-1994 (R2002)**

**Northeast Metrology Corp.**

4490 Broadway  
Depew, NY 14043  
Basile Korbut 716-827-3770  
nem@nemcal.com

**CALIBRATION, DIMENSIONAL MEASUREMENT, AND TESTING**

Valid to: **June 6, 2024**

Certificate Number: **ACT-1116**

**CALIBRATION**

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source <sup>1,2</sup>	Up to 330 mV (0.33 to 3.3) V (3.3 to 33) V (33 to 330) V (330 to 1 020) V	7.1 $\mu$ V 40 $\mu$ V 0.4 mV 5.9 mV 20 mV	Multiproduct Calibrator
DC Voltage – Measure <sup>1,2</sup>	Up to 200 mV (0.2 to 2) V (2 to 20) V (20 to 200) V (200 to 1 020) V	7.6 $\mu$ V 37 $\mu$ V 0.32 mV 7 mV 55 mV	Keithley 2002 8.5 Digit Multimeter
DC Current – Source <sup>1,2</sup>	Up to 330 $\mu$ A (0.33 to 3.3) mA (3.3 to 33) mA (33 to 330) mA (0.33 to 1.1) A (1.1 to 11) A (11 to 20.5) A	72 nA 0.38 $\mu$ A 3.5 $\mu$ A 35 $\mu$ A 0.61 mA 5.1 mA 24 mA	Multiproduct Calibrator
DC Current – Source <sup>1,2</sup> Clamp-On Ammeters	(20.5 to 205) A	208 mA	Multiproduct Calibrator w/ 10-turn Coil
	(205 to 1 025) A	1 A	Multiproduct Calibrator w/ 10-turn Coil



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Measure <sup>1,2</sup>	Up to 200 $\mu$ A	0.1 $\mu$ A	Keithley 2002 8.5 Digit Multimeter
	(0.2 to 2) mA	0.99 $\mu$ A	
	(2 to 20) mA	10 $\mu$ A	
	(20 to 200) mA	0.13 mA	
	(0.2 to 2) A	2.2 mA	
DC Current – Measure <sup>1,2</sup>	(2 to 10) A	16 mA	Fluke 8846A 6.5 Digit Multimeter
	(10 to 400) A	8.2 A	Fluke 8846A 6.5 Digit Multimeter w/ Current Shunt
	(400 to 2 000) A	19 A	Fluke 353 Clamp Meter
AC Voltage – Source <sup>1,2</sup>	Up to 33 mV		Multiproduct Calibrator
	10 Hz to 10 kHz	16 $\mu$ V	
	(10 to 20) kHz	15 $\mu$ V	
	(20 to 50) kHz	40 $\mu$ V	
	(50 to 100) kHz	0.13 mV	
	(100 to 500) kHz	0.3 mV	
	(33 to 330) mV		
	(10 to 45) Hz	0.13 mV	
	45 Hz to 10 kHz	0.13 mV	
	(10 to 20) kHz	73 $\mu$ V	
	(20 to 50) kHz	0.13 mV	
	(50 to 100) kHz	0.28 mV	
	(100 to 500) kHz	0.74 mV	
	(0.33 to 3.3) V		
	(10 to 45) Hz	1.3 mV	
	45 Hz to 10 kHz	1.3 mV	
	(10 to 20) kHz	0.72 mV	
(20 to 50) kHz	1.1 mV		
(50 to 100) kHz	2.3 mV		
(100 to 500) kHz	8.9 mV		
(3.3 to 33) V			
(10 to 45) Hz	9.1 mV		
45 Hz to 10 kHz	14 mV		
(10 to 20) kHz	9.6 mV		
(20 to 50) kHz	13 mV		
(50 to 100) kHz	32 mV		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source <sup>1,2</sup>	(33 to 330) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (330 V to 1 020) V 45 Hz to 5 kHz (5 to 10) kHz	0.11 V 73 mV 97 mV 0.11 V 0.74 V 0.26 V 0.32 V	Multiproduct Calibrator
AC Voltage – Measure <sup>1,2</sup>	Up to 200 mV (20 to 50) Hz (50 to 100) Hz 100 Hz to 10 kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz (100 to 200) kHz 200 kHz to 2 MHz (0.22 to 2) V (20 to 100) Hz 100 Hz to 10 kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz (100 to 200) kHz 200 kHz to 2 MHz (2 to 20) V (20 to 50) Hz 50 Hz to 1 kHz (1 to 5) kHz (5 to 25) kHz (25 to 50) kHz (50 to 100) kHz (20 to 200) V (20 to 50) Hz 50 Hz to 1 kHz (2 to 10) kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz	1.6 mV 0.21 mV 99 μV 0.11 mV 0.76 mV 1.9 μV 5.1 mV 13 mV 2.1 mV 1 mV 1.1 mV 7.6 mV 19 mV 51 mV 0.13 V 37 mV 11 mV 13 mV 16 mV 76 mV 0.19 V 0.38 V 0.12 V 0.14 V 0.17 V 0.77 V 1.9 V	Keithley 2002 8.5 Digit Multimeter

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1,2</sup>	(200 to 750) V (20 to 50) Hz 50 Hz to 1 kHz (2 to 10) kHz (10 to 25) kHz	1.5 V 1.1 V 1.2 V 1.4 V	Keithley 2002 8.5 Digit Multimeter
AC Current – Source <sup>1,2</sup>	(29 to 330) $\mu$ A (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz (0.33 to 3.3) mA (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz (3.3 to 33) mA (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz (33 to 330) mA (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz (0.33 to 1.1) A (10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (1.1 to 3) A (10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.52 $\mu$ A 0.52 $\mu$ A 0.87 $\mu$ A 2.6 $\mu$ A 3.9 $\mu$ A 3.2 $\mu$ A 3.2 $\mu$ A 5.2 $\mu$ A 16 $\mu$ A 18 $\mu$ A 16 $\mu$ A 16 $\mu$ A 28 $\mu$ A 68 $\mu$ A 0.13 mA 0.63 mA 0.16 mA 0.16 mA 0.4 mA 0.75 mA 1.5 mA 0.69 mA 0.69 mA 7.8 mA 33 mA 1.9 mA 1.9 mA 19 mA 78 mA	Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source <sup>1,2</sup>	(3 to 11) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	8.7 mA 12 mA 0.3 A	Multiproduct Calibrator
	(11 to 20.5) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	31 mA 26 mA 0.62 A	
AC Current – Source <sup>1,2</sup> Clamp-On Ammeters	(20.5 to 205) A (45 to 440) Hz	0.37 A	Multiproduct Calibrator, 10-turn Coil
	(205 to 1 025) A (45 to 440) Hz	1.5 A	Multiproduct Calibrator, 50-turn Coil
Resistance – Source <sup>1,2</sup>	Up to 11 Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω 330 Ω to 1.1 kΩ (1.1 to 3.3) kΩ (3.3 to 11) kΩ (11 to 33) kΩ (33 to 110) kΩ (110 to 330) kΩ 330 kΩ to 1.1 MΩ (1.1 to 3.3) MΩ (3.3 to 11) MΩ (11 to 33) MΩ (33 to 110) MΩ (110 to 330) MΩ 330 MΩ to 1.1 GΩ	1.7 mΩ 2.5 mΩ 4.8 mΩ 11 mΩ 33 mΩ 0.11 Ω 0.32 Ω 1.1 Ω 3.2 Ω 12 Ω 37 Ω 0.51 kΩ 0.5 kΩ 0.97 kΩ 62 kΩ 21 kΩ 6.4 MΩ	Multiproduct Calibrator
Resistance – Measure <sup>1,2</sup>	Up to 20 Ω (20 to 200) Ω (0.2 to 2) kΩ (2 to 20) kΩ (20 to 200) kΩ (0.2 to 2) MΩ (2 to 20) MΩ (20 to 200) MΩ (0.2 to 1) GΩ	1.8 mΩ 8.2 mΩ 41 mΩ 0.44 Ω 6.5 Ω 0.11 kΩ 4.3 kΩ 1.9 MΩ 18 MΩ	Keithley 2002 8.5 Digit Multimeter



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Source <sup>1,2</sup>	(0.19 to 3.3) nF	14 pF	Multiproduct Calibrator
	(3.3 to 11) nF	40 pF	
	(11 to 33) nF	0.13 nF	
	(110 to 330) nF	1.1 nF	
	(0.33 to 1.1) μF	4 nF	
	(1.1 to 3.3) μF	13 nF	
	(3.3 to 11) μF	40 nF	
	(11 to 33) μF	0.16 μF	
	(33 to 110) μF	0.64 μF	
	(110 to 330) μF	2.5 μF	
	(0.33 to 1.1) mF	8.1 μF	
	(1.1 to 3.3) mF	25 μF	
	(3.3 to 11) mF	57 μF	
	(11 to 33) mF	0.51 mF	
(33 to 110) mF	1.4 mF		
Electrical Simulation of Thermocouple Indicating Devices <sup>1</sup>	Type E		Multiproduct Calibrator
	(-250 to -100) °C	0.88 °C	
	(-100 to -25) °C	0.74 °C	
	(-25 to 350) °C	0.73 °C	
	(350 to 650) °C	0.74 °C	
	(650 to 1 000) °C	0.75 °C	
	Type J		
	(-210 to -100) °C	0.66 °C	
	(-100 to -30) °C	0.62 °C	
	(-30 to 150) °C	0.62 °C	
	(150 to 760) °C	0.62 °C	
	(760 to 1 200) °C	0.64 °C	
	Type K		
	(-200 to -100) °C	0.68 °C	
	(-100 to -25) °C	0.63 °C	
	(-25 to 120) °C	0.62 °C	
	(120 to 1 000) °C	0.65 °C	
	(1 000 to 1 372) °C	0.72 °C	
	Type R		
	(0 to 250) °C	3.64 °C	
(250 to 400) °C	3.62 °C		
(400 to 1 000) °C	3.62 °C		
(1 000 to 1 767) °C	3.62 °C		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices <sup>1</sup>	Type S		Multiproduct Calibrator
	(0 to 250) °C	3.63 °C	
	(250 to 1 000) °C	3.62 °C	
	(1 000 to 1 400) °C	3.62 °C	
	(1 400 to 1 767) °C	3.63 °C	
	Type T		
	(-250 to -150) °C	0.87 °C	
	(150 to 0) °C	0.65 °C	
(0 to 120) °C	0.62 °C		
(120 to 400) °C	0.62 °C		

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gage Blocks	(0.05 to 1) in (1 to 4) in (4 to 20) in	3.3 μin 5.4 μin 9.9 μin	Grade 0 & 00 Gage Blocks, Gage Block Comparator
Bench Micrometer	Up to 1 in (lead screw)	6.3 μin	Grade 0 Gage Blocks, Precision Ball, Oz Testers
Outside Micrometer <sup>1</sup>	Up to 24 in	602 μin	Grade 0 Gage Blocks
Height Gage <sup>1</sup>	Up to 38 in	79 μin	Grade 0 Gage Blocks
Height Master <sup>1</sup>	Up to 12 in	35 μin	Grade 0 Gage Blocks, Digital Indicator
Plain Plug Gage <sup>1</sup>	Up to 8 in	16 μin	Grade 0 Gage Blocks, P&W Supermicrometer
Plain Ring Gage <sup>1</sup>	(0.04 to 12) in	25 μin	Grade 0 Gage Blocks, Ring Gage Comparator
Thread Plug <sup>1</sup>	(0.06 to 10) in	28 μin	Grade 0 Gage Blocks, P&W Supermicrometer, Thread Wires
Thread Ring <sup>1</sup>	Up to 10 in	28 μin	Set Plugs, Bore Gage
Tapered Pipe Thread Plug <sup>1</sup>	Up to 4 in	38 μin	Thread Wires, Sine Block, Supermicrometer, Height Master, Indicator

### Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Tapered Sine Block <sup>1</sup>	(0.062 5 to 3) in	15 μin	3/8 Ball, P&W Supermicrometer, Gage Block
Thread Measuring Wire <sup>1</sup>	(4 to 100) TPI	6.1 μin	Class X Cylindrical Roll, P&W Supermicrometer
Micrometer Standard <sup>1</sup>	(1 to 20) in	49 μin	Grade 0 Gage Blocks, P&W Supermicrometer
Caliper <sup>1</sup>	Up to 60 in	0.001 in	Grade 0 Gage Blocks
Test Indicator <sup>1</sup>	Up to 0.06 in	60 μin	Grade 0 Gage Blocks
Plunger Indicator <sup>1</sup>	Up to 2 in	31 μin	Grade 0 Gage Blocks
Optical Comparator <sup>1</sup> Squareness	Up to 18 in	91 μin	Master Square
Linearity	Up to 18 in	107 μin	Optical Magnification Checker, Gage Blocks
Magnification	10X, 20X, 31.25X, 50X, 100X	487 μin	Glass Scale
Surface Plates <sup>1,3</sup> Grades AA, A and B Overall Flatness	Up to 180 in x 240 in	(25 + 0.7DL) μin	Precision Level System
Local Area Flatness (repeat reading)	Up to 0.001 6 in	15.7 μin	Repeat-O-Meter
Squares, Steps, Angle Plates <sup>1</sup>	Up to 18 in	62 μin	Squareness Checker
Angles & Angle Blocks <sup>1</sup>	Up to 90 °	81 μin	Electronic Gage Head, Mu Checker, Sine Bar
Spheres <sup>1</sup>	Up to 1 in	11 μin	Grade 0 Gage Blocks, P&W Supermicrometer
Measuring Tapes	Up to 30 ft	0.06 in	Lixer Master, Lecia Laser
Rulers	Up to 12 in (12 to 72) in	73 μin 0.06 in	Optical Comparator Leica Laser



**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Radius Gages (Leaf Style)	(0.0625 to 2) in	85 $\mu$ in	OGP Smart Scope
Thickness / Feeler Gages	(0.001 to .2) in	9.2 $\mu$ in	Grade 0 Gage Blocks, P&W Supermicrometer

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure – Measure <sup>1</sup>	Up to 1 000 psig	3 psi	Heise Gauge
Pressure – Source <sup>1</sup>	Up to 200 psig Up to 1 500 psig Up to 15 000 psig	0.025 psi 3.1 psi 31 psi	Ametek Calibrator Mansfield & Green Deadweight Tester
Vacuum – Source <sup>1</sup>	(-30 to 0) inHg	0.017 inHg	Manometer
Weights	Up to 20 g Up to 32 000 g	0.17 mg 4.2 g	ASTM E617 Class 1 Weights, Mass Comparator
Scales and Balances <sup>1</sup>	Up to 20 g	30 $\mu$ g	ASTM E617 Class 1 Weights and internal calibration procedure utilized in the calibration of the weighing system.
Scales <sup>1</sup>	Up to 1 000 lb	1.3 lb	NIST Class F Weights and internal calibration procedure utilized in the calibration of the weighing system.
Durometers (Force Only) Type A & D	(0 to 90) Duro	0.6 Duro	Durocalibrator



ANSI National Accreditation Board

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Rockwell Hardness Testers <sup>1,4</sup>	(40 to 95) HR15N	0.19 HR15N	Indirect Verification per ASTM E18 using Hardness Test Blocks.
	(20 to 79) HR15TW	0.37 HR15TW	
	≥ 80 HR15TW	0.21 HR15TW	
	≥ 80 HR15W	0.67 HR15W	
	≥80 HR15X	0.33 HR15X	
	(85 to 93) HR15Y	1.3 HR15Y	
	≥94 HR15Y	0.63 HR15Y	
	(40 to 59) HR30N	0.55 HR30N	
	(60 to 85) HR30N	0.28 HR30N	
	(20 to 49) HR30TW	0.9 HR30TW	
	(50 to 56) HR30TW	0.66 HR30TW	
	≥57 HR30TW	0.39 HR30TW	
	(40 to 64) HR30W	0.9 HR30W	
	≥65 HR30W	0.76 HR30W	
	(60 to 78) HR30X	0.99 HR30X	
	≥79 HR30X	0.15 HR30X	
	(60 to 87) HR30Y	0.82 HR30Y	
	≥88 HR30Y	0.37 HR30Y	
(10 to 49) HR45N	0.43 HR45N		
(50 to 66) HR45N	0.22 HR45N		
(67 to 75) HR45N	0.19 HR45N		
(1 to 39) HR45TW	0.73 HR45TW		
≥40 HR45TW	0.41 HR45TW		
(10 to 47) HR45W	0.3 HR45W		
≥48 HR45W	0.13 HR45W		
(40 to 68) HR45X	0.81 HR45X		
≥69 HR45X	0.35 HR45X		



ANSI National Accreditation Board

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Rockwell Hardness Testers <sup>1,4</sup>	(60 to 81) HR45Y	0.94 HR45Y	Indirect Verification per ASTM E18 using Hardness Test Blocks.
	≥82 HR45Y	0.35 HR45Y	
	(60 to 69) HRA	0.28 HRA	
	(70 to 79) HRA	0.17 HRA	
	(80 to 85) HRA	0.16 HRA	
	Carbide ≥86 HRA	0.16 HRA	
	(1 to 50) HRBW	1.4 HRBW	
	(51 to 79) HRBW	0.87 HRBW	
	≥80 HRBW	0.42 HRBW	
	(20 to 39) HRC	0.4 HRC	
	(40 to 59) HRC	0.36 HRC	
	(60 to 70) HRC	0.32 HRC	
	(40 to 49) HRD	0.27 HRD	
	(50 to 69) HRD	0.26 HRD	
	(70 to 80) HRD	0.18 HRD	
	≥65 HREW	0.54 HREW	
	(40 to 69) HRFW	0.54 HRFW	
	≥70 HRFW	0.4 HRFW	
	(1 to 39) HRGW	0.76 HRGW	
	≥40 HRGW	0.3 HRGW	
(60 to 79) HRHW	0.54 HRHW		
≥80 HRHW	0.41 HRHW		
(10 to 29) HRKW	0.64 HRKW		
≥30 HRKW	0.40 HRKW		
≥90 HRLW	0.36 HRLW		
≥70 HRMW	0.56 HRMW		
(40 to 84) HRP	0.91 HRP		
≥85 HRRW	0.65 HRRW		

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Rockwell Hardness Testers <sup>1,4</sup>	(100 to 119) HRSW	0.41 HRSW	Indirect Verification per ASTM E18 using Hardness Test Blocks.
	≥120 HRSW	0.24 HRSW	
	(110 to 111) HRVW	0.95 HRVW	
	≥112 HRVW ≥80 HRVW	0.2 HRVW 0.21 HRVW	
Vickers Hardness Testers <sup>1</sup>	HV1		Indirect Verification per ASTM E384 using Hardness Test Blocks.
	200 HV	8.7 HV	
	400 HV	21 HV	
	700 HV	44 HV	
	HV2		
	200 HV	6.9 HV	
	400 HV	16 HV	
	700 HV	31 HV	
	HV5		
	200 HV	3.9 HV	
	400 HV	11 HV	
	700 HV	20 HV	
	HV10		
	200 HV	3.1 HV	
	400 HV	7.7 HV	
	700 HV	15 HV	
	HV20		
	200 HV	2.5 HV	
	400 HV	6.2 HV	
	700 HV	11 HV	
	HV30		
	200 HV	2 HV	
	400 HV	4.4 HV	
	700 HV	9.3 HV	
HV50			
200 HV	1.9 HV		
400 HV	3.5 HV		
700 HV	6.3 HV		
HV 0.01			
400 HV	30 HV		
700 HV	40 HV		
HV 0.025			
200 HV	9 HV		
400 HV	20 HV		
700 HV	30 HV		

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Vickers Hardness Testers <sup>1</sup>	HV 0.05		Indirect Verification per ASTM E384 using Hardness Test Blocks.
	200 HV	8.5 HV	
	400 HV	19 HV	
	700 HV	27 HV	
	HV 0.1		
	200 HV	8 HV	
	400 HV	18 HV	
	700 HV	25 HV	
	HV 0.2		
	200 HV	7 HV	
	400 HV	17 HV	
	700 HV	20 HV	
	HV 0.3		
	200 HV	6 HV	
400 HV	16 HV		
700 HV	19 HV		
HV 0.5			
200 HV	5 HV		
400 HV	15 HV		
700 HV	17 HV		
Brinell Hardness Testers <sup>1</sup>	HBW 1 / 62.5		Indirect Verification per ASTM E10 using Hardness Test Blocks.
	(200 to 400) HBW	2 HBW	
	(400 to 600) HBW	4 HBW	
	HBW 2.5 / 187.5		
	(200 to 400) HBW	2 HBW	
	(400 to 600) HBW	4 HBW	
	HBW 10 / 500		
	(200 to 400) HBW	2 HBW	
	(400 to 600) HBW	4 HBW	
	HBW 5 / 1 000		
	(200 to 400) HBW	2 HBW	
	(400 to 600) HBW	4 HBW	
	HBW 10 / 1 000		
	(200 to 400) HBW	2 HBW	
	(400 to 600) HBW	4 HBW	
	HBW 10 / 1500		
(200 to 400) HBW	2 HBW		
(400 to 600) HBW	4 HBW		
HBW 10 / 2 000			
(200 to 400) HBW	2 HBW		
(400 to 600) HBW	4 HBW		

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Brinell Hardness Testers <sup>1</sup>	HBW 10 / 2 500 (200 to 400) HBW (400 to 600) HBW HBW 10 / 3 000 (200 to 400) HBW (400 to 600) HBW	2 HBW 4 HBW 2 HBW 4 HBW	Indirect Verification per ASTM E10 using Hardness Test Blocks.
Knoop Hardness Testers <sup>1</sup>	HK 0.01 400 HK 700 HK HK 0.025 200 HK 400 HK 700 HK HK 0.05 200 HK 400 HK 700 HK HK 0.1 200 HK 400 HK 700 HK HK 0.2 200 HK 400 HK 700 HK HK 0.3 200 HK 400 HK 700 HK HK 0.5 200 HK 400 HK 700 HK HK 1 200 HK 400 HK 700 HK	30 HK 40 HK 9 HK 20 HK 30 HK 8.5 HK 19 HK 27 HK 8 HK 18 HK 25 HK 7 HK 17 HK 20 HK 6 HK 16 HK 19 HK 5 HK 15 HK 17 HK 5 HK 10 HK 15 HK	Indirect Verification per ASTM E384 using Hardness Test Blocks.
Torque Tools	(25 to 250) lbf·in (25 to 250) lbf·ft (100 to 1 000) lbf·ft	0.66 % of reading 0.61 % of reading 0.61 % of reading	Norbar Torque Calibrator

### Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Force – Compression and Tension <sup>1</sup>	(0 to 10) lbf	0.12 % of reading + 0.25 lbf	NIST Class F Weights
	(0 to 100) lbf	0.12 % of reading + 0.5 lbf	
	(100 to 1 124) lbf	0.06 % of reading	Load Cells
	(1 000 to 3 000) lbf	0.05 % of reading	
	(3 000 to 11 240) lbf	0.05 % of reading	
(10 000 to 20 000) lbf	0.002 % of reading		
(20 000 to 50 000) lbf	0.002 % of reading		
Force – Compression	(50 000 to 300 000) lbf	0.1 % of reading	Load Cells

### Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature – Measure <sup>1</sup>	(-35 to 200) °C	0.08 °C	Fluke 1524 Indicator, Fluke 5627A PRT
Humidity – Measure <sup>1</sup>	(0 to 90) % RH	1.1 % RH	MI70 / HMP77 Temperature/Humidity Indicator

### Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Measure <sup>1</sup>	(1.1 to 10) Hz	9.5 mHz	Keithley 2002 8.5 Digit Multimeter
	(10 to 100) Hz	0.58 Hz	
	100 Hz to 1.1 kHz	0.37 Hz	
	(1.1 to 10) kHz	58 mHz	
	(10 to 100) kHz	36 Hz	
	100 kHz to 1.1 MHz	0.37 kHz	
	(1.1 to 10) MHz	0.58 kHz	
	(10 to 15) MHz	6 kHz	
Tachometers (Non-Contact Type)	(30 to 500 000) rpm	2.5 rpm	Comparison to Nova-Strobe PBL Stroboscope
Stopwatches & Timers (Push Button)	Up to 24 hr	55 ms	Frequency Counter

**Time and Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electric Timers (Pulse)	1 ms to 24 hr	4.5 $\mu$ s	Frequency Counter

**DIMENSIONAL MEASUREMENT**

**3 Dimensional**

Specific Tests and / or Properties Measured	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
3D Dimensional Parts	X = (0 to 25) in Y = (0 to 35) in Z = (0 to 17) in	272 $\mu$ in 293 $\mu$ in 264 $\mu$ in	B&S Bridge Type CMM, Customer Drawing, PC DMIS S/W

**TESTING**

**Mechanical**

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Comments
Hardness – Rockwell and Superficial Scales	ASTM E18	Sintered and Wrought Products	Wilson Rockwell Hardness Tester, Rockwell Scales – A, B, C, E, F, G, & H Superficial Scales – T & N
Hardness – Brinell	ASTM E10	Sintered and Wrought Products	Newage Dyna Brinell HB500 to HB3 000
Hardness – Portable	ASTM E110	Sintered and Wrought Products	Mitutoyo Portable
Hardness – Leeb, Equotip	ASTM A956	Sintered and Wrought Products	Equotip, LD Scale
Microhardness – Vickers	ASTM E92	Sintered and Wrought Products	Mitutoyo MVK, Heavy Load Vickers 500 g to 5 kg



**Mechanical**

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Comments
Microhardness – Knoop and Vickers	ASTM E384 ASTM B933	Sintered and Wrought Products	Buehler Micro Hardness Tester, Knoop Scale – (10 to 1 000) g Range Vickers Scale – (10 to 1 000) g Range

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. CMC for Electrical – DC/Low Frequency Calibrations do not include possible contributions to uncertainty from a “best available” unit under test.
3. The use of (DL) signifies the Diagonal Length of the surface plate in feet.
4. The following statement is from the NIST recommended practice guide (Special Publication 960-5, page 60, section 8.3.3, paragraph 2) “Currently, there are no generally agreed upon U.S. or international methods for calculating the measurement uncertainty of a Rockwell hardness machine or the uncertainty in the certified value of standardized test blocks.”
5. This scope is formatted as part of a single document including Certificate of Accreditation No. ACT-1116.



R. Douglas Leonard Jr., VP, PILR SBU