

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

Northeast Metrology Corp.

4490 Broadway

Depew, NY 14043

Basile Korbut 716-827-3770

nem@nemcal.com

CALIBRATION, DIMENSIONAL MEASUREMENT, AND TESTING

Valid to: **June 6, 2026**

Certificate Number: **ACT-1116**

CALIBRATION

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source ^{1,2}	Up to 330 mV (0.33 to 3.3) V (3.3 to 33) V (33 to 330) V (330 to 1 020) V	7.1 μV 40 μV 0.4 mV 5.9 mV 20 mV	Multiproduct Calibrator
DC Voltage – Measure ^{1,2}	Up to 200 mV (0.2 to 2) V (2 to 20) V (20 to 200) V (200 to 1 020) V	7.6 μV 37 μV 0.32 mV 7 mV 55 mV	8.5 Digit Multimeter
DC Current – Source ^{1,2}	Up to 330 μA (0.33 to 3.3) mA (3.3 to 33) mA (33 to 330) mA (0.33 to 1.1) A (1.1 to 11) A (11 to 20.5) A	72 nA 0.38 μA 3.5 μA 35 μA 0.61 mA 5.1 mA 24 mA	Multiproduct Calibrator
DC Current – Source ^{1,2} (Clamp-On Ammeters)	(20.5 to 205) A	0.21 A	Multiproduct Calibrator with 10-turn Coil
DC Current – Source ^{1,2} (Clamp-On Ammeters)	(205 to 1 025) A	1 A	Multiproduct Calibrator with 50-turn Coil

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Measure ^{1,2}	Up to 200 μ A (0.2 to 2) mA (2 to 20) mA (20 to 200) mA (0.2 to 2) A	0.1 μ A 0.99 μ A 10 μ A 0.13 mA 2.2 mA	8.5 Digit Multimeter
DC Current – Measure ^{1,2}	(2 to 10) A	16 mA	6.5 Digit Multimeter
DC Current – Measure ^{1,2}	(10 to 400) A	8.2 A	6.5 Digit Multimeter with Current Shunt
DC Current – Measure ^{1,2}	(400 to 2 000) A	19 A	Clamp-on Meter
AC Voltage – Source ^{1,2}	Up to 33 mV 10 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (33 to 330) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (0.33 to 3.3) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (3.3 to 33) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	16 μ V 15 μ V 40 μ V 0.13 mV 0.3 mV 0.13 mV 0.13 mV 73 μ V 0.13 mV 0.28 mV 0.74 mV 1.3 mV 1.3 mV 0.72 mV 1.1 mV 2.3 mV 8.9 mV 9.1 mV 14 mV 9.6 mV 13 mV 32 mV	Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ^{1,2}	(33 to 330) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (330 V to 1 020) V 45 Hz to 5 kHz (5 to 10) kHz	0.11 V 73 mV 97 mV 0.11 V 0.74 V 0.26 V 0.32 V	Multiproduct Calibrator
AC Voltage – Measure ^{1,2}	Up to 200 mV (20 to 50) Hz (50 to 100) Hz 100 Hz to 10 kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz (100 to 200) kHz 200 kHz to 2 MHz (0.22 to 2) V (20 to 100) Hz 100 Hz to 10 kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz (100 to 200) kHz 200 kHz to 2 MHz (2 to 20) V (20 to 50) Hz 50 Hz to 1 kHz (1 to 5) kHz (5 to 25) kHz (25 to 50) kHz (50 to 100) kHz (20 to 200) V (20 to 50) Hz 50 Hz to 1 kHz (2 to 10) kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz	1.6 mV 0.21 mV 99 μV 0.11 mV 0.76 mV 1.9 μV 5.1 mV 13 mV 2.1 mV 1 mV 1.1 mV 7.6 mV 19 mV 51 mV 0.13 V 37 mV 11 mV 13 mV 16 mV 76 mV 0.19 V 0.38 V 0.12 V 0.14 V 0.17 V 0.77 V 1.9 V	8.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure ^{1,2}	(200 to 750) V		8.5 Digit Multimeter
	(20 to 50) Hz	1.5 V	
	50 Hz to 1 kHz	1.1 V	
	(2 to 10) kHz	1.2 V	
	(10 to 25) kHz	1.4 V	
AC Current – Source ^{1,2}	(29 to 330) μ A		Multiproduct Calibrator
	(20 to 45) Hz	0.52 μ A	
	45 Hz to 1 kHz	0.52 μ A	
	(1 to 5) kHz	0.87 μ A	
	(5 to 10) kHz	2.6 μ A	
	(10 to 30) kHz	3.9 μ A	
	(0.33 to 3.3) mA		
	(20 to 45) Hz	3.2 μ A	
	45 Hz to 1 kHz	3.2 μ A	
	(1 to 5) kHz	5.2 μ A	
	(5 to 10) kHz	16 μ A	
	(10 to 30) kHz	18 μ A	
	(3.3 to 33) mA		
	(20 to 45) Hz	16 μ A	
	45 Hz to 1 kHz	16 μ A	
	(1 to 5) kHz	28 μ A	
	(5 to 10) kHz	68 μ A	
	(10 to 30) kHz		
	(33 to 330) mA	0.13 mA	
	(10 to 20) Hz	0.63 mA	
	(20 to 45) Hz	0.16 mA	
	45 Hz to 1 kHz	0.16 mA	
	(1 to 5) kHz	0.4 mA	
	(5 to 10) kHz	0.75 mA	
	(10 to 30) kHz	1.5 mA	
	(0.33 to 1.1) A		
	(10 to 45) Hz	0.69 mA	
	45 Hz to 1 kHz	0.69 mA	
	(1 to 5) kHz	7.8 mA	
	(5 to 10) kHz	33 mA	
(1.1 to 3) A			
(10 to 45) Hz	1.9 mA		
45 Hz to 1 kHz	1.9 mA		
(1 to 5) kHz	19 mA		
(5 to 10) kHz	78 mA		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source ^{1,2}	(3 to 11) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	8.7 mA 12 mA 0.3 A	Multiproduct Calibrator
	(11 to 20.5) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	31 mA 26 mA 0.62 A	
AC Current – Source ^{1,2} (Clamp-On Ammeters)	(20.5 to 205) A (45 to 440) Hz	0.37 A	Multiproduct Calibrator with 10-turn Coil
AC Current – Source ^{1,2} (Clamp-On Ammeters)	(205 to 1 025) A (45 to 440) Hz	1.5 A	Multiproduct Calibrator with 50-turn Coil
Resistance – Source ^{1,2} (Simulation)	Up to 11 Ω	1.7 mΩ	Multiproduct Calibrator
	(11 to 33) Ω	2.5 mΩ	
	(33 to 110) Ω	4.8 mΩ	
	(110 to 330) Ω	11 mΩ	
	330 Ω to 1.1 kΩ	33 mΩ	
	(1.1 to 3.3) kΩ	0.11 Ω	
	(3.3 to 11) kΩ	0.32 Ω	
	(11 to 33) kΩ	1.1 Ω	
	(33 to 110) kΩ	3.2 Ω	
	(110 to 330) kΩ	12 Ω	
	330 kΩ to 1.1 MΩ	37 Ω	
	(1.1 to 3.3) MΩ	0.51 kΩ	
	(3.3 to 11) MΩ	0.5 kΩ	
(11 to 33) MΩ	0.97 kΩ		
(33 to 110) MΩ	62 kΩ		
(110 to 330) MΩ	21 kΩ		
330 MΩ to 1.1 GΩ	6.4 MΩ		
Resistance – Measure ^{1,2}	Up to 20 Ω	1.8 mΩ	8.5 Digit Multimeter
	(20 to 200) Ω	8.2 mΩ	
	(0.2 to 2) kΩ	41 mΩ	
	(2 to 20) kΩ	0.44 Ω	
	(20 to 200) kΩ	6.5 Ω	
	(0.2 to 2) MΩ	0.11 kΩ	
	(2 to 20) MΩ	4.3 kΩ	
	(20 to 200) MΩ	1.9 MΩ	
(0.2 to 1) GΩ	18 MΩ		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Source ^{1,2} (Simulation)	(0.19 to 3.3) nF (3.3 to 11) nF (11 to 33) nF (110 to 330) nF (0.33 to 1.1) μF (1.1 to 3.3) μF (3.3 to 11) μF (11 to 33) μF (33 to 110) μF (110 to 330) μF (0.33 to 1.1) mF (1.1 to 3.3) mF (3.3 to 11) mF (11 to 33) mF (33 to 110) mF	14 pF 40 pF 0.13 nF 1.1 nF 4 nF 13 nF 40 nF 0.16 μF 0.64 μF 2.5 μF 8.1 μF 25 μF 57 μF 0.51 mF 1.4 mF	Multiproduct Calibrator
Electrical Simulation of Thermocouple Indicating Devices ¹	Type E (-250 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1 000) °C Type J (-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1 200) °C Type K (-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1 000) °C (1 000 to 1 372) °C Type R (0 to 250) °C (250 to 400) °C (400 to 1 000) °C (1 000 to 1 767) °C	0.88 °C 0.74 °C 0.73 °C 0.74 °C 0.75 °C 0.66 °C 0.62 °C 0.62 °C 0.62 °C 0.64 °C 0.68 °C 0.63 °C 0.62 °C 0.65 °C 0.72 °C 3.64 °C 3.62 °C 3.62 °C 3.62 °C	Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices ¹	Type S		Multiproduct Calibrator
	(0 to 250) °C	3.63 °C	
	(250 to 1 000) °C	3.62 °C	
	(1 000 to 1 400) °C	3.62 °C	
	(1 400 to 1 767) °C	3.63 °C	
	Type T		
	(-250 to -150) °C	0.87 °C	
	(150 to 0) °C	0.65 °C	
(0 to 120) °C	0.62 °C		
(120 to 400) °C	0.62 °C		

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gage Blocks ³	Up to 0.09 in (0.1 to 4) in (> 4 to 20) in	4 µin (2.6 + 1.8L) µin (0.5 + 2.1L) µin	Grade 0 & 00 Gage Blocks, Gage Block Comparator
1D Measuring Machine ¹ (lead screw)	Up to 1 in	8.6 µin	Grade 0 Gage Blocks, Precision Ball, Oz Testers
Outside Micrometer ^{1,3} (5 µin resolution) (50 µin resolution) (100 µin resolution) (0.001 in resolution)	Up to 1 in Up to 6 in Up to 40 in Up to 40 in	8.8 µin (48 + 1.1L) µin (81 + 1.8L) µin (604 + 0.7L) µin	Grade 0 Gage Blocks
Height Gage ^{1,3} (0.000 004 in resolution) (0.000 01 in resolution) (0.000 05 in resolution) (0.000 1 in resolution) (0.000 5 in resolution) (0.001 in resolution)	Up to 38 in Up to 43 in Up to 43 in Up to 60 in Up to 60 in Up to 60 in	(4.4 + 2.6L) µin (6.6 + 2.6L) µin (38 + 2.2L) µin (72 + 2L) µin (391 + 6.2L) µin (702 + 4.6L) µin	Grade 0 Gage Blocks
Height Master ¹	Up to 12 in	65 µin	Grade 0 Gage Blocks, Digital Indicator
Plain Plug Gage ^{1,3}	Up to 8 in	(15 + 2D) µin	Grade 0 Gage Blocks, P&W Supermicrometer®
Plain Ring Gage ^{1,3}	(0.04 to 12) in	(7.5 + 3.6D) µin	Grade 0 Gage Blocks, Ring Gage Comparator

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thread Plug ^{1,3}	(0.06 to 6) in	$(58 + 3.4D) \mu\text{in}$	Grade 0 Gage Blocks, P&W Supermicrometer [®] , Thread Wires
Thread Ring ^{1,3} Pitch Diameter	Up to 3 in	$(58 + 3.4D) \mu\text{in}$	Thread Setting Plug
Minor Diameter	(0.04 to 0.2) in (0.2 to 3) in	$(15 + 2D) \mu\text{in}$ $(50 + 7.2D) \mu\text{in}$	Pin Gages, Bore Micrometer
Tapered Pipe Thread Plug ^{1,3}	Up to 3 in	$(134 + 1.5D) \mu\text{in}$	Thread Wires, Sine Block, P & W Supermicrometer [®] , Height Master, Indicator
Tapered Sine Block ¹	(0.062 5 to 3) in	71 μin	3/8 Gage Ball, P&W Supermicrometer [®] , Gage Blocks
Thread Measuring Wire ¹	(4 to 100) TPI	15.5 μin	Class X Cylindrical Roll, P&W Supermicrometer [®]
Micrometer Standard ^{1,3}	Up to 6 in (6 to 20) in	$(35 + 1.9L) \mu\text{in}$ $(35 + 2.4L) \mu\text{in}$	Grade 0 Gage Blocks, P&W Supermicrometer [®]
Calipers ^{1,3} (5 μin resolution) (0.001 in resolution)	Up to 60 in Up to 60 in	$(360 + 3.5L) \mu\text{in}$ $(550 + 16.7L) \mu\text{in}$	Grade 0 Gage Blocks
Test Indicator ¹ (50 μin resolution) (100 μin resolution) (500 μin resolution) (0.001 in resolution)	Up to 0.06 in Up to 0.06 in Up to 0.06 in Up to 0.06 in	51 μin 95 μin 474 μin 726 μin	Grade 0 Gage Blocks
Plunger Indicator ^{1,3} (50 μin resolution) (100 μin resolution) (0.001 in resolution)	Up to 2 in Up to 2 in Up to 2 in	$(44 + 6.4L) \mu\text{in}$ $(72 + 2.2L) \mu\text{in}$ $(704 + 22L) \mu\text{in}$	Grade 0 Gage Blocks
Surface Plates ^{1,3} Grades AA, A and B Overall Flatness	Up to 159 inDL	$(25 + 0.7DL) \mu\text{in}$	In accordance with ASME B89.3.7 using Precision Level System
Local Area Flatness (repeat reading)	Up to 0.001 6 in	15.7 μin	Repeat-O-Meter
Squares, Steps, Angle Plates ¹	Up to 18 in	167 μin	Squareness Checker

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Spheres ¹	Up to 1 in	36 μin	Grade 0 Gage Blocks, P&W Supermicrometer [®]
Rulers	Up to 6 in	70 μin	OGP Smart Scope
Thickness / Feeler Gages	(0.001 to 0.2) in	13.9 μin	Grade 0 Gage Blocks, P&W Supermicrometer [®]

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure – Measure ¹	Up to 1 000 psig	3 psi	Comparison to Heise Gauge
Pressure – Source ¹	Up to 200 psig Up to 1 500 psig Up to 15 000 psig	0.025 psi 3.1 psi 31 psi	Pressure Calibrator, Deadweight Tester
Vacuum – Source ¹	(-30 to 0) inHg	0.017 inHg	Manometer
Scales and Balances ¹	Up to 20 g	30 μg	ASTM E617 Class 1 Weights and internal calibration procedure utilized in the calibration of the weighing system.
Scales ¹	Up to 1 000 lb	1.3 lb	NIST Class F Weights and internal calibration procedure utilized in the calibration of the weighing system.
Durometers Types A, D Spring Force Only	(0 to 90) Duro	0.6 Duro	Partial Verification per ASTM D2240 using Durocalibrator

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Rockwell Hardness Testers ^{1,4}	(40 to 95) HR15N	0.19 HR15N	Indirect Verification per ASTM E18 using Hardness Test Blocks.
	(20 to 79) HR15TW ≥ 80 HR15TW	0.37 HR15TW 0.21 HR15TW	
	≥ 80 HR15W	0.67 HR15W	
	≥ 80 HR15X	0.33 HR15X	
	(85 to 93) HR15Y ≥ 94 HR15Y	1.3 HR15Y 0.63 HR15Y	
	(40 to 59) HR30N (60 to 85) HR30N	0.55 HR30N 0.28 HR30N	
	(20 to 49) HR30TW (50 to 56) HR30TW ≥ 57 HR30TW	0.9 HR30TW 0.66 HR30TW 0.39 HR30TW	
	(40 to 64) HR30W ≥ 65 HR30W	0.9 HR30W 0.76 HR30W	
	(60 to 78) HR30X ≥ 79 HR30X	0.99 HR30X 0.15 HR30X	
	(60 to 87) HR30Y ≥ 88 HR30Y	0.82 HR30Y 0.37 HR30Y	
	(10 to 49) HR45N (50 to 66) HR45N (67 to 75) HR45N	0.43 HR45N 0.22 HR45N 0.19 HR45N	
	(1 to 39) HR45TW ≥ 40 HR45TW	0.73 HR45TW 0.41 HR45TW	
	(10 to 47) HR45W ≥ 48 HR45W	0.3 HR45W 0.13 HR45W	
	(40 to 68) HR45X ≥ 69 HR45X	0.81 HR45X 0.35 HR45X	



ANSI National Accreditation Board

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Rockwell Hardness Testers ^{1,4}	(60 to 81) HR45Y	0.94 HR45Y	Indirect Verification per ASTM E18 using Hardness Test Blocks.
	≥ 82 HR45Y	0.35 HR45Y	
	(60 to 69) HRA	0.28 HRA	
	(70 to 79) HRA	0.17 HRA	
	(80 to 85) HRA	0.16 HRA	
	Carbide ≥ 86 HRA	0.16 HRA	
	(1 to 50) HRBW	1.4 HRBW	
	(51 to 79) HRBW	0.87 HRBW	
	≥ 80 HRBW	0.42 HRBW	
	(20 to 39) HRC	0.4 HRC	
	(40 to 59) HRC	0.36 HRC	
	(60 to 70) HRC	0.32 HRC	
	(40 to 49) HRD	0.27 HRD	
	(50 to 69) HRD	0.26 HRD	
	(70 to 80) HRD	0.18 HRD	
	≥ 65 HREW	0.54 HREW	
	(40 to 69) HRFW	0.54 HRFW	
	≥ 70 HRFW	0.4 HRFW	
	(1 to 39) HRGW	0.76 HRGW	
	≥ 40 HRGW	0.3 HRGW	
(60 to 79) HRHW	0.54 HRHW		
≥ 80 HRHW	0.41 HRHW		
(10 to 29) HRKW	0.64 HRKW		
≥ 30 HRKW	0.40 HRKW		
≥90 HRLW	0.36 HRLW		
≥70 HRMW	0.56 HRMW		
(40 to 84) HRP	0.91 HRP		
≥85 HRRW	0.65 HRRW		

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Rockwell Hardness Testers ^{1,4}	(100 to 119) HRSW ≥ 120 HRSW (110 to 111) HRVW ≥ 112 HRVW ≥ 80 HRVW	0.41 HRSW 0.24 HRSW 0.95 HRVW 0.2 HRVW 0.21 HRVW	Indirect Verification per ASTM E18 using Hardness Test Blocks.
Vickers Hardness Testers ¹	HV1 200 HV 400 HV 700 HV HV2 200 HV 400 HV 700 HV HV5 200 HV 400 HV 700 HV HV10 200 HV 400 HV 700 HV HV20 200 HV 400 HV 700 HV HV30 200 HV 400 HV 700 HV HV50 200 HV 400 HV 700 HV HV 0.01 400 HV 700 HV HV 0.025 200 HV 400 HV 700 HV	8.7 HV 21 HV 44 HV 6.9 HV 16 HV 31 HV 3.9 HV 11 HV 20 HV 3.1 HV 7.7 HV 15 HV 2.5 HV 6.2 HV 11 HV 2 HV 4.4 HV 9.3 HV 1.9 HV 3.5 HV 6.3 HV 30 HV 40 HV 9 HV 20 HV 30 HV	Indirect Verification per ASTM E384 using Hardness Test Blocks.

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Vickers Hardness Testers ¹	HV 0.05 200 HV 400 HV 700 HV HV 0.1 200 HV 400 HV 700 HV HV 0.2 200 HV 400 HV 700 HV HV 0.3 200 HV 400 HV 700 HV HV 0.5 200 HV 400 HV 700 HV	8.5 HV 19 HV 27 HV 8 HV 18 HV 25 HV 7 HV 17 HV 20 HV 6 HV 16 HV 19 HV 5 HV 15 HV 17 HV	Indirect Verification per ASTM E384 using Hardness Test Blocks.
Brinell Hardness Testers ¹	HBW 1 / 62.5 (200 to 400) HBW (400 to 600) HBW HBW 2.5 / 187.5 (200 to 400) HBW (400 to 600) HBW HBW 10 / 500 (200 to 400) HBW (400 to 600) HBW HBW 5 / 1 000 (200 to 400) HBW (400 to 600) HBW HBW 10 / 1 000 (200 to 400) HBW (400 to 600) HBW HBW 10 / 1500 (200 to 400) HBW (400 to 600) HBW HBW 10 / 2 000 (200 to 400) HBW (400 to 600) HBW	2 HBW 4 HBW 2 HBW 4 HBW 2 HBW 4 HBW 2 HBW 4 HBW 2 HBW 4 HBW 2 HBW 4 HBW 2 HBW 4 HBW	Indirect Verification per ASTM E10 using Hardness Test Blocks.

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Brinell Hardness Testers ¹	HBW 10 / 2 500 (200 to 400) HBW (400 to 600) HBW HBW 10 / 3 000 (200 to 400) HBW (400 to 600) HBW	2 HBW 4 HBW 2 HBW 4 HBW	Indirect Verification per ASTM E10 using Hardness Test Blocks.
Knoop Hardness Testers ¹	HK 0.01 400 HK 700 HK HK 0.025 200 HK 400 HK 700 HK HK 0.05 200 HK 400 HK 700 HK HK 0.1 200 HK 400 HK 700 HK HK 0.2 200 HK 400 HK 700 HK HK 0.3 200 HK 400 HK 700 HK HK 0.5 200 HK 400 HK 700 HK HK 1 200 HK 400 HK 700 HK	30 HK 40 HK 9 HK 20 HK 30 HK 8.5 HK 19 HK 27 HK 8 HK 18 HK 25 HK 7 HK 17 HK 20 HK 6 HK 16 HK 19 HK 5 HK 15 HK 17 HK 5 HK 10 HK 15 HK	Indirect Verification per ASTM E384 using Hardness Test Blocks.
Torque Tools	(25 to 250) lbf·in (25 to 250) lbf·ft (100 to 1 000) lbf·ft	0.68 % of reading 0.67 % of reading 0.68 % of reading	Norbar Torque Calibrator



ANSI National Accreditation Board

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Force – Compression and Tension ¹	(> 0 to 10) lbf (10 to 100) lbf	0.12 % of reading + 0.25 lbf 0.12 % of reading + 0.5 lbf	NIST Class F Weights
Force – Compression and Tension ¹	(100 to 1 124) lbf (1 000 to 3 000) lbf (3 000 to 11 240) lbf (10 000 to 20 000) lbf (20 000 to 50 000) lbf	0.06 % of reading 0.05 % of reading 0.05 % of reading 0.002 % of reading 0.002 % of reading	Load Cells
Force – Compression (Only)	(50 000 to 300 000) lbf	0.1 % of reading	Load Cells

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature – Measure ¹	(-35 to 200) °C	0.08 °C	Reference Thermometer with PRT
Humidity – Measure ¹	(0 to 90) %RH	1.1 %RH	Temperature/Humidity Indicator

Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Measure ¹	(1.1 to 10) Hz (10 to 100) Hz 100 Hz to 1.1 kHz (1.1 to 10) kHz (10 to 100) kHz 100 kHz to 1.1 MHz (1.1 to 10) MHz (10 to 15) MHz	9.5 mHz 0.58 Hz 0.37 Hz 58 mHz 36 Hz 0.37 kHz 0.58 kHz 6 kHz	8.5 Digit Multimeter
Tachometers ³ (Non-Contact Type)	(30 to 500 000) rpm	2.5 rpm	Comparison to Nova-Strobe PBL Stroboscope
Stopwatches & Timers (Push Button)	Up to 24 hr	55 ms	Comparison to Bench Timer
Electric Timers (Pulse)	1 ms to 24 hr	4.5 μs	Comparison to Frequency Counter

DIMENSIONAL MEASUREMENT

3 Dimensional

Specific Tests and / or Properties Measured	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
3D Dimensional Parts	X-axis = Up to 25 in Y-axis = Up to 35 in Z-axis = Up to 17 in	267 μ in 325 μ in 232 μ in	Coordinate Measuring Machine; Customer Drawings, PC DMIS Software

TESTING

Mechanical

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Comments
Hardness – Rockwell and Superficial Scales	ASTM E18	Sintered and Wrought Products	Wilson Rockwell Hardness Tester, Rockwell Scales – A, B, C, E, F, G, & H Superficial Scales – T & N
Hardness – Brinell	ASTM E10	Sintered and Wrought Products	Newage Dyna Brinell HB500 to HB3 000
Hardness – Portable	ASTM E110	Sintered and Wrought Products	Mitutoyo Portable
Hardness – Leeb, Equotip	ASTM A956	Sintered and Wrought Products	Equotip, LD Scale
Microhardness – Vickers	ASTM E92	Sintered and Wrought Products	Mitutoyo MVK, Heavy Load Vickers (500 to 5 000) g
Microhardness – Knoop and Vickers	ASTM E384 ASTM B933	Sintered and Wrought Products	Buehler Micro Hardness Tester, Knoop Scale – (10 to 1 000) g Range Vickers Scale – (10 to 1 000) g Range

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. CMC for Electrical – DC/Low Frequency Calibrations do not include possible contributions to uncertainty from a “best available” unit under test.
3. L = length in inches; D = diameter in inches; DL = diagonal length in inches; rpm = revolutions per minute.
4. The following statement is from the NIST recommended practice guide (Special Publication 960-5, page 60, section 8.3.3, paragraph 2) “Currently, there are no generally agreed upon U.S. or international methods for calculating the measurement uncertainty of a Rockwell hardness machine or the uncertainty in the certified value of standardized test blocks.”
5. This scope is formatted as part of a single document including Certificate of Accreditation No. ACT-1116.



Jason Stine, Vice President

